

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 01/17/2023

OPERATOR NAME:  
JEREMY HINMAN  
Permit #: 39036  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:14
INTERNAL STANDARD	VERIFIED	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18

*OK / Lr 1/17/23*

