

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 02/18/2024

OPERATOR NAME:
OFFICER HINMAN
Permit #: 39036
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:36
INTERNAL STANDARD	VERIFIED	12:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:37
BLANK TEST	0.00	12:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:39
BLANK TEST	0.00	12:40

