

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 02/28/2024

OPERATOR NAME: GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:58
INTERNAL STANDARD	VERIFIED	12:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:59
BLANK TEST	0.00	13:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:00
BLANK TEST	0.00	13:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:01
BLANK TEST	0.00	13:02

