

SOLUTION CHANGE

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 03/07/2023

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:57
INTERNAL STANDARD	VERIFIED	13:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	13:57
BLANK TEST	0.00	13:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:59
BLANK TEST	0.00	14:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:00
BLANK TEST	0.00	14:01

