

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 03/13/2023

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:43
INTERNAL STANDARD	VERIFIED	07:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:45
BLANK TEST	0.00	07:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:46
BLANK TEST	0.00	07:47

