

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 03/25/2024

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:56
INTERNAL STANDARD	VERIFIED	07:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:57
BLANK TEST	0.00	07:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:58
BLANK TEST	0.00	07:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:59
BLANK TEST	0.00	08:00

