

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 04/15/2024

OPERATOR NAME: GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:58
INTERNAL STANDARD	VERIFIED	09:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:01
BLANK TEST	0.00	10:02

