

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 04/17/2023

OPERATOR NAME:
 GREGORY D GILFUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:49
INTERNAL STANDARD	VERIFIED	07:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:49
BLANK TEST	0.00	07:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:50
BLANK TEST	0.00	07:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:52
BLANK TEST	0.00	07:53

