

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 04/24/2023

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:25
INTERNAL STANDARD	VERIFIED	14:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:25
BLANK TEST	0.00	14:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:27
BLANK TEST	0.00	14:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:28
BLANK TEST	0.00	14:29

