

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 05/06/2024

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:24
INTERNAL STANDARD	VERIFIED	12:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:25
BLANK TEST	0.00	12:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:26
BLANK TEST	0.00	12:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:28
BLANK TEST	0.00	12:28

