

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 06/19/2023

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:06
INTERNAL STANDARD	VERIFIED	07:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:06
BLANK TEST	0.00	07:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:09
BLANK TEST	0.00	07:10

