

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 06/20/2022

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:41
INTERNAL STANDARD	VERIFIED	07:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:44
BLANK TEST	0.00	07:45

