

SUPERVISOR TEST

AUBURN CITY  
 POLICE DEPT  
 DATAMASTER dmt: 105406

Date: 07/24/2023

OPERATOR NAME:  
 GREGORY D GILFUS *RG*  
 Permit #: 30728  
 Operator Agency/Dept: AUBURN CITY PD  
 Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:34
INTERNAL STANDARD	VERIFIED	07:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:35
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:37
BLANK TEST	0.00	07:38

