

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 07/25/2022

OPERATOR NAME: GREGORY D GILFUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:37
INTERNAL STANDARD	VERIFIED	07:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:39
BLANK TEST	0.00	07:40
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:40
BLANK TEST	0.00	07:41

