

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 07/31/2023

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:25
INTERNAL STANDARD	VERIFIED	07:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	07:25
BLANK TEST	0.00	07:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:28
BLANK TEST	0.00	07:29

