

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 08/08/2022

OPERATOR NAME:
JEREMY HINMAN
Permit #: 39036
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:33
INTERNAL STANDARD	VERIFIED	12:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:33
BLANK TEST	0.00	12:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:35
BLANK TEST	0.00	12:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	12:36
BLANK TEST	0.00	12:37

Oct 17 14 0438

