

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 08/29/2022

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:27
INTERNAL STANDARD	VERIFIED	07:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:28
BLANK TEST	0.00	07:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:29
BLANK TEST	0.00	07:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:30
BLANK TEST	0.00	07:31

