

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 09/11/2023

OPERATOR NAME:
 GREGORY D GILFUS
 Permit #: 30726
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:18
INTERNAL STANDARD	VERIFIED	07:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:19
BLANK TEST	0.00	07:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:21
BLANK TEST	0.00	07:22

