

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 09/12/2022

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:39
INTERNAL STANDARD	VERIFIED	07:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:40
BLANK TEST	0.00	07:41
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:41
BLANK TEST	0.00	07:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:43

