

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 09/17/2023

OPERATOR NAME:
JEREMY HINMAN
Permit #: 39036
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:31
INTERNAL STANDARD	VERIFIED	12:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:31
BLANK TEST	0.00	12:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:32
BLANK TEST	0.00	12:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:34
BLANK TEST	0.00	12:35

OK JH #0438 09/17/23

