

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 09/26/2022

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:14
INTERNAL STANDARD	VERIFIED	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:16
BLANK TEST	0.00	08:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:17
BLANK TEST	0.00	08:18

