

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 10/09/2023

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:26
INTERNAL STANDARD	VERIFIED	07:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:27
BLANK TEST	0.00	07:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:28
BLANK TEST	0.00	07:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:29
BLANK TEST	0.00	07:30

