

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 10/16/2023

OPERATOR NAME:  
OFFICER HINMAN  
Permit #: 39036  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:54
INTERNAL STANDARD	VERIFIED	09:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:55
BLANK TEST	0.00	09:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:56
BLANK TEST	0.00	09:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:57
BLANK TEST	0.00	09:58

