

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 10/31/2022

OPERATOR NAME: *[Signature]*  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:45
INTERNAL STANDARD	VERIFIED	07:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:45
BLANK TEST	0.00	07:46
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:47
BLANK TEST	0.00	07:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:48
BLANK TEST	0.00	07:49

