

SUPERVISOR TEST

AUBURN CITY  
 POLICE DEPT  
 DATAMASTER dmt: 105406

Date: 11/07/2022

OPERATOR NAME: GREGORY D GILFUS  
 Permit #: 30728  
 Operator Agency/Dept: AUBURN CITY PD  
 Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:36
INTERNAL STANDARD	VERIFIED	07:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:37
BLANK TEST	0.00	07:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:39
BLANK TEST	0.00	07:40

