

# SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 11/21/2022

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:50
INTERNAL STANDARD	VERIFIED	07:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:51
BLANK TEST	0.00	07:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:52
BLANK TEST	0.00	07:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	07:53
BLANK TEST	0.00	07:54

