

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 11/21/2023

OPERATOR NAME: GREGORY D GILFUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:34
INTERNAL STANDARD	VERIFIED	07:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:36
BLANK TEST	0.00	07:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:37
BLANK TEST	0.00	07:38

