

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 11/28/2022

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:26
INTERNAL STANDARD	VERIFIED	08:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:27
BLANK TEST	0.00	08:27
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:28
BLANK TEST	0.00	08:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	08:29
BLANK TEST	0.00	08:30

