

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 12/04/2023

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:30
INTERNAL STANDARD	VERIFIED	07:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:30
BLANK TEST	0.00	07:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:32
BLANK TEST	0.00	07:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:33
BLANK TEST	0.00	07:34

