

SOLUTION CHANGE

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 03/01/2022

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:33
INTERNAL STANDARD	VERIFIED	11:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:33
BLANK TEST	0.00	11:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:34
BLANK TEST	0.00	11:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:36
BLANK TEST	0.00	11:36

