

SOLUTION CHANGE

AUBURN CITY  
 POLICE DEPT  
 DATAMASTER dmt: 105406

Date: 09/05/2022

OPERATOR NAME:  
 GREGORY D GILFUS  
 Permit #: 30728  
 Operator Agency/Dept: AUBURN CITY PD  
 Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:48
INTERNAL STANDARD	VERIFIED	10:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:50
BLANK TEST	0.00	10:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:51
BLANK TEST	0.00	10:52

