

SOLUTION CHANGE

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 12/01/2022

OPERATOR NAME: GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:53
INTERNAL STANDARD	VERIFIED	14:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:53
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:55
BLANK TEST	0.00	14:56
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:56
BLANK TEST	0.00	14:57

