

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 01/18/2021

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 20430

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 06:30 |
| INTERNAL STANDARD       | VERIFIED | 06:30 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 06:31 |
| BLANK TEST              | 0.00     | 06:32 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 06:32 |
| BLANK TEST              | 0.00     | 06:33 |
| REFERENCE STANDARD TEMP | 34.00c   |       |
| REFERENCE STANDARD      | 0.09     | 06:33 |
| BLANK TEST              | 0.00     | 06:34 |

