

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 01/25/2021

OPERATOR NAME:
 GREGORY D GILFUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:35
INTERNAL STANDARD	VERIFIED	08:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:37
BLANK TEST	0.00	08:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:38
BLANK TEST	0.00	08:39

