

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 02/07/2022

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:18
INTERNAL STANDARD	VERIFIED	08:18
REFERENCE STANDARD TEMP	54.00c	
REFERENCE STANDARD	0.10	08:19
BLANK TEST	0.00	08:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:20
BLANK TEST	0.00	08:21
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:21
BLANK TEST	0.00	08:22

