

SUPERVISOR TEST

AUBURN CITY  
 POLICE DEPT  
 DATAMASTER dmt: 105406

Date: 02/15/2021

OPERATOR NAME:  
 GREGORY D GILFUS  
 Permit #: 30728  
 Operator Agency/Dept: AUBURN CITY PD  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:07
INTERNAL STANDARD	VERIFIED	09:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:08
BLANK TEST	0.00	09:09
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:09
BLANK TEST	0.00	09:10
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:10
BLANK TEST	0.00	09:11

