

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 02/22/2021

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:41
INTERNAL STANDARD	VERIFIED	08:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:42
BLANK TEST	0.00	08:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:44
BLANK TEST	0.00	08:45

