

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

te: 03/15/2021

OPERATOR NAME: *[Signature]*
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:33
INTERNAL STANDARD	VERIFIED	10:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:34
BLANK TEST	0.00	10:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:35
BLANK TEST	0.00	10:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:36
BLANK TEST	0.00	10:37

