

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 03/22/2021

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:47
INTERNAL STANDARD	VERIFIED	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:47
BLANK TEST	0.00	08:48
REFERENCE STANDARD TEMP	34.86c	
REFERENCE STANDARD	0.09	08:48
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:50
BLANK TEST	0.00	08:51

