

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 04/04/2022

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:29
INTERNAL STANDARD	VERIFIED	08:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:30
BLANK TEST	0.00	08:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:31
BLANK TEST	0.00	08:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:32
BLANK TEST	0.00	08:33

