

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 04/05/2021

OPERATOR NAME:  
GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:37
INTERNAL STANDARD	VERIFIED	08:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:37
BLANK TEST	0.00	08:38
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:39
BLANK TEST	0.00	08:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:40
BLANK TEST	0.00	08:41

