

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 04/19/2021

OPERATOR NAME: Gregory D Gilfus
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:49
INTERNAL STANDARD	VERIFIED	08:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:49
BLANK TEST	0.00	08:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:50
BLANK TEST	0.00	08:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:52
BLANK TEST	0.00	08:52

