

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 04/26/2021

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:05
INTERNAL STANDARD	VERIFIED	10:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:06
BLANK TEST	0.00	10:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:07
BLANK TEST	0.00	10:08
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:08
BLANK TEST	0.00	10:09

