

SUPERVISOR TEST

AUBURN CITY  
POLICE DEPT  
DATAMASTER dmt: 105406

Date: 04/26/2022

OPERATOR NAME: GREGORY D GILFUS  
Permit #: 30728  
Operator Agency/Dept: AUBURN CITY PD  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:54
INTERNAL STANDARD	VERIFIED	07:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:54
BLANK TEST	0.00	07:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:56
BLANK TEST	0.00	07:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	07:57
BLANK TEST	0.00	07:58

