

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 05/18/2020

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:28
INTERNAL STANDARD	VERIFIED	08:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:29
BLANK TEST	0.00	08:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:30
BLANK TEST	0.00	08:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:31
BLANK TEST	0.00	08:32

