

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 05/24/2021

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21020

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:54
INTERNAL STANDARD	VERIFIED	13:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:54
BLANK TEST	0.00	13:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:55
BLANK TEST	0.00	13:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:57
BLANK TEST	0.00	13:57

