

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 05/24/2022

OPERATOR NAME:
JEREMY HINMAN
Permit #: 39036
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:12
INTERNAL STANDARD	VERIFIED	08:12
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:12
BLANK TEST	0.00	08:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:14
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16

OC. J. H. #043P

