

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 06/08/2020

OPERATOR NAME: GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 20179

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:04
INTERNAL STANDARD	VERIFIED	12:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:04
BLANK TEST	0.00	12:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:05
BLANK TEST	0.00	12:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:06
BLANK TEST	0.00	12:07

