

SUPERVISOR TEST

AUBURN CITY
 POLICE DEPT
 DATAMASTER dmt: 105406

Date: 06/28/2021

OPERATOR NAME: *GREGORY D GILEUS*
 GREGORY D GILEUS
 Permit #: 30728
 Operator Agency/Dept: AUBURN CITY PD
 Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:58
INTERNAL STANDARD	VERIFIED	07:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	07:59
BLANK TEST	0.00	07:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:00
BLANK TEST	0.00	08:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:01
BLANK TEST	0.00	08:02

