

SUPERVISOR TEST

AUBURN CITY
POLICE DEPT
DATAMASTER dmt: 105406

Date: 07/05/2021

OPERATOR NAME:
GREGORY D GILFUS
Permit #: 30728
Operator Agency/Dept: AUBURN CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:00
INTERNAL STANDARD	VERIFIED	08:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:00
BLANK TEST	0.00	08:01
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:02
BLANK TEST	0.00	08:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:03
BLANK TEST	0.00	08:04

